



Neilson Research Corporation  
245 S Grape St  
Medford, OR 97501  
TEL: (541) 770-5678 FAX: (541) 770-2901  
Website: www.nrclabs.com

September 16, 2019

Andy Chasteen  
Medford School District  
815 S Oakdale Ave  
Medford, OR 97501  
TEL: 541-842-3646  
FAX: 541-842-1160

RE: Lone Pine Elementary Lead Testing

Order No.: 19090343

Dear Andy Chasteen:

Neilson Research Corporation received 2 sample(s) on 9/9/2019 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,  
Neilson Research Corporation

Tamra Schmedemann  
Senior Project Manager  
245 S Grape St  
Medford, OR 97501

Original



Neilson Research Corporation  
245 S Grape St  
Medford, OR 97501  
TEL: (541) 770-5678 FAX: (541) 770-2901  
Website: www.nrclabs.com

## Case Narrative

WO#: 19090343  
Date: 9/16/2019

---

**CLIENT:** Medford School District  
**Project:** Lone Pine Elementary Lead Testing

---

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

---

Original



Neilson Research Corporation  
 245 S Grape St  
 Medford, OR 97501  
 TEL: (541) 770-5678 FAX: (541) 770-2901  
 Website: www.nrclabs.com

# Analytical Report

WO#: 19090343  
 Date Reported: 9/16/2019

Medford School District  
 815 S Oakdale Ave  
 Medford , OR 97501

**Lab Order:** 19090343  
**Received Date:** 9/9/2019 12:38:00 PM  
**Reported Date:** 9/16/2019 3:40:23 PM

Sample Information:

**Lab ID:** 19090343-01      Client Sample ID: Cafeteria Drinking Fountain  
 Collection Date: 9/9/2019 9:11:00 AM      Collected By:  
 Matrix: Drinking Water      Sample Location: Cafeteria Drinking Fountain

Trace Metals by EPA 200.8 ICP-MS								Analyst: SJS
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	0.00115		0.000100	mg/L	1	9/10/2019	0.0150	A

**Lab ID:** 19090343-02      Client Sample ID: Gym Lobby Drinking Fountain  
 Collection Date: 9/9/2019 9:40:00 AM      Collected By:  
 Matrix: Drinking Water      Sample Location: Gym Lobby Drinking Fountain

Trace Metals by EPA 200.8 ICP-MS								Analyst: SJS
Analyses	Result	Qual	MRL	Units	DF	Date Analyzed	MCL	NELAP Status
Lead	0.00130		0.000100	mg/L	1	9/10/2019	0.0150	A

<b>QUALIFIERS</b>	* Value exceeds Maximum Contaminant Level.	C1 Sample container temperature is out of limit as specified at testcode
	E Value above quantitation range	H Holding times for preparation or analysis exceeded
	J Analyte detected below quantitation limits	MI Recovery outside control limits due to Matrix Interference
	ND Not Detected at the Reporting Limit	PL Permit Limit
	R RPD outside accepted recovery limits	

Original



Neilson Research Corporation  
 245 S Grape St  
 Medford, OR 97501  
 TEL: (541) 770-5678 FAX: (541) 770-2901  
 Website: www.nrclabs.com

# QC SUMMARY REPORT

WO#: 19090343  
 16-Sep-19

**Client:** Medford School District  
**Project:** Lone Pine Elementary Lead Testing

**TestCode:** LEAD\_DW

Sample ID: <b>MB-1725</b>	SampType: <b>MBLK</b>	TestCode: <b>LEAD_DW</b>	Units: <b>mg/L</b>	Prep Date: <b>9/10/2019</b>	RunNo: <b>4007</b>						
Client ID: <b>PBW</b>	Batch ID: <b>1725</b>	TestNo: <b>E200.8</b>	<b>E200.8</b>	Analysis Date: <b>9/10/2019</b>	SeqNo: <b>78739</b>						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead	ND	0.000100									

Sample ID: <b>LCS-1725</b>	SampType: <b>LCS</b>	TestCode: <b>LEAD_DW</b>	Units: <b>mg/L</b>	Prep Date: <b>9/10/2019</b>	RunNo: <b>4007</b>						
Client ID: <b>LCSW</b>	Batch ID: <b>1725</b>	TestNo: <b>E200.8</b>	<b>E200.8</b>	Analysis Date: <b>9/10/2019</b>	SeqNo: <b>78740</b>						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead	0.104	0.000100	0.1000	0	104	85	115				

Sample ID: <b>19090344-01AMS</b>	SampType: <b>MS</b>	TestCode: <b>LEAD_DW</b>	Units: <b>mg/L</b>	Prep Date: <b>9/10/2019</b>	RunNo: <b>4007</b>						
Client ID: <b>BatchQC</b>	Batch ID: <b>1725</b>	TestNo: <b>E200.8</b>	<b>E200.8</b>	Analysis Date: <b>9/10/2019</b>	SeqNo: <b>78762</b>						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead	0.101	0.000100	0.1000	0.0002140	101	70	130				

Sample ID: <b>19090344-01AMSD</b>	SampType: <b>MSD</b>	TestCode: <b>LEAD_DW</b>	Units: <b>mg/L</b>	Prep Date: <b>9/10/2019</b>	RunNo: <b>4007</b>						
Client ID: <b>BatchQC</b>	Batch ID: <b>1725</b>	TestNo: <b>E200.8</b>	<b>E200.8</b>	Analysis Date: <b>9/10/2019</b>	SeqNo: <b>78763</b>						
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Lead	0.102	0.000100	0.1000	0.0002140	102	70	130	0.1014	0.700	20	

**Qualifiers:** C1 Sample container temperature is out of limit as specified at testcode      H Holding times for preparation or analysis exceeded      MI Recovery outside control limits due to Matrix In  
 ND Not Detected at the Reporting Limit      PL Permit Limit      RL Reporting Detection Limit

Original

